




NOTES/SPECIFICATIONS:

1. INPUT BEAM DIAMETER: 2 - 9 mm
2. OUTPUT BEAM SIZE: 2.9 x 4.0 mm² - 11.4 x 16.0 mm²
3. WORKING DISTANCE: 40 mm
4. OPERATING WAVELENGTH RANGE: 350 - 1600 nm (UNCOATED)
5. INCLUDED BUT NOT SHOWN:
   0.05" HEX KEY FOR LOCKING SCREWS;
   SM1 INTERNAL THREAD DUST CAPS SM1CP1 (2X)

FOR INFORMATION ONLY
NOT FOR MANUFACTURING PURPOSES

DRAWING PROJECTION		
	NAME	DATE
DRAWN	KS	31/MAY/22
APPROVAL	CS	31/MAY/22

COPYRIGHT © 2022 BY THORLABS

VALUES IN PARENTHESIS ARE CALCULATED AND MAY CONTAIN ROUNDOFF ERRORS

THORLABS
www.thorlabs.com

Fly's Eye Homogenizer

MATERIAL	REV
N/A	A

ITEM #	APPROX WEIGHT
FLE1	0.08 kg